

Tibial Nail

Tibial Nail

Specification	
D	L
Φ8.0	260
Φ8.0	280
Φ8.0	300
Φ8.0	320
Φ8.0	340
Φ8.0	360
Φ8.5	260
Φ8.5	280
Φ8.5	300
Φ8.5	320
Φ8.5	340
Φ8.5	360
Φ9.0	260
Φ9.0	280
Φ9.0	300
Φ9.0	320
Φ9.0	340
Φ9.0	360
Φ10	260
Φ10	280
Φ10	300
Φ10	320
Φ10	340
Φ10	360

Tibial Locking Screws–B

Specification	
D	L
Φ4.2	18
Φ4.2	20
Φ4.2	22
Φ4.2	24
Φ4.2	26
Φ4.2	28
Φ4.2	30
Φ4.2	32
Φ4.2	34
Φ4.2	36
Φ4.2	38
Φ4.2	40

Tibial Condyles Locking Screws

Specification	
D	L
Φ5.0	31
Φ5.0	35
Φ5.0	39
Φ5.0	43
Φ5.0	47
Φ5.0	51
Φ5.0	55
Φ5.0	59
Φ5.0	63
Φ5.0	67
Φ5.0	71
Φ5.0	75
Φ5.0	79
Φ5.0	83
Φ5.0	87
Φ5.0	91
Φ5.0	95

Tibial Locking Screws–B

Specification	
D	L
Φ4.2	42
Φ4.2	44
Φ4.2	46
Φ4.2	48
Φ4.2	50
Φ4.2	52
Φ4.2	54
Φ4.2	56
Φ4.2	58
Φ4.2	60
Φ4.2	62
Φ4.2	64
Φ4.2	66
Φ4.2	68
Φ4.2	70
Φ4.2	72
Φ4.2	74
Φ4.2	76
Φ4.2	78
Φ4.2	80

Tibial Locking Screws–A

Specification	
D	L
Φ4.2	18
Φ4.2	20
Φ4.2	22
Φ4.2	24
Φ4.2	26
Φ4.2	28
Φ4.2	30
Φ4.2	32
Φ4.2	34
Φ4.2	36
Φ4.2	38
Φ4.2	40
Φ4.2	42
Φ4.2	44
Φ4.2	46
Φ4.2	48
Φ4.2	50
Φ4.2	52
Φ4.2	54
Φ4.2	56
Φ4.2	58
Φ4.2	60
Φ4.2	62
Φ4.2	64
Φ4.2	66
Φ4.2	68
Φ4.2	70
Φ4.2	72
Φ4.2	74
Φ4.2	76
Φ4.2	78
Φ5.0	24
Φ5.0	26
Φ5.0	28
Φ5.0	30
Φ5.0	32
Φ5.0	34
Φ5.0	36
Φ5.0	38

Tibial Locking Screws–A

Specification	
D	L
Φ5.0	40
Φ5.0	42
Φ5.0	44
Φ5.0	46
Φ5.0	48
Φ5.0	50
Φ5.0	52
Φ5.0	54
Φ5.0	56
Φ5.0	58
Φ5.0	60
Φ5.0	62
Φ5.0	64
Φ5.0	66
Φ5.0	70
Φ5.0	72
Φ5.0	74
Φ5.0	76
Φ5.0	78
Φ5.0	80
Φ5.0	85
Φ5.0	90
Φ5.0	95

Tibial End Caps

L
0
5
10
15

Tibial Nail

Features:

1. Proximal



---Medio-lateral locking design enables compression and dynamization;

2. Distal



---Oblique locking design to prevent soft tissue damage;

---Medio-lateral and antero-posterior locking options for stability of distal fragments;

3. End cap



---Prevents ingrowth of tissue and facilitates nail extraction;

---Creates a fixed-angle construct by end cap connection with the most proximal screw;



4. Locking Screws

---Dual thread design for locking screw to increase purchase and stability;

5. Our products are manufactured using one or more of the following materials: commercially pure titanium(ISO 5832-2, ASTM F67),Ti6Al4V alloy(ISO 5832-3, ASTM F1472).

Indications:

1. Tibial diaphyseal fractures;
2. Tibial metaphysis fractures;
3. Part of intra-articular fractures for tibia plateau and PILON fractures;